

EU DECLARATION OF CONFORMITY

FOR MEDICAL DEVICES AND PERSONAL PROTECTIVE EQUIPMENT

Originator: J.F ROBLES

Revision: 013

Revision date: 13.07.2020

Validity date: 27.02.2025

PRODUCT	SHIELDskin™ ORANGE NITRILE™ 260
DESCRIPTION	Powder-free Extra Length Ambidextrous Non-
	Sterile 26 cm Nitrile Gloves
CLASSIFICATION	Medical Device Class 1 / Personal Protective
	Equipment (PPE) Category III (Complex Design)

SHIELD Scientific codes	Sizes
67 6231	6/XS
67 6232	7/S
67 6233	8/M
67 6234	9/L
67 6235	10/XL

The manufacturer established in the Union:

SHIELD Scientific B.V.

Dr Willem Dreeslaan 1 – 6721 ND BENNEKOM – THE NETHERLANDS

declares under his/her sole responsibility that the Medical Device and PPE (product codes as mentioned above) described hereafter:

SHIELDskin[™] ORANGE NITRILE[™] 260

is in conformity with the provisions of the Medical Device Regulation (EU) 2017/745 and with the national standards transposing harmonized standards EN 455-1:2000, EN 455-2:2015, EN 455-3:2015 and EN 455-4:2009. It is selfcertified as a Medical Device Class 1.

is in conformity with the provisions of Regulation (EU) 2016/425 and with the harmonized standards ISO 374-1:2016 (as a Type B glove against reagents: J, K, P & T), ISO 374-2:2019 (performance level 3, including protection against viruses), EN 16523-1:2015 + A1:2018, ISO 374-4:2019, ISO 374-5:2016 and ISO 21420:2020. This device is identical to the PPE, which is the subject of EU Type Examination (Module B) certificate of conformity no. FI18/962117 issued by the Notified Body:

SGS FIMKO OY (Notified Body No: 0598) Takomotie 8, FI-00380 Helsinki, Finland

This device is subject to the procedure set out in Article VIII (Module D) of the Regulation under the surveillance of the Notified Body:

SGS FIMKO OY (Notified Body No: 0598) Takomotie 8, FI-00380 Helsinki, Finland

Signed for and behalf of SHIELD Scientific B.V

Date: 13th July 2020 Place: Bennekom Validity of this declaration: 13th July 2020 until 27th February 2025

J.F ROBLES **General Manager**